

Limulus Amebocyte Lysate (LAL) – Report of Analysis

Test Performed	Endosafe - PTS LAL Assay
Test Date	02/17/2026
Analyst	VICTOR
Instrument	Endosafe® nexgen-PTS™ (SN: 25201216)

Sample Information

Company Name	Novera Compounds
Sample Tested	BPC-157
Sample Lot Number	395715
Sample Concentration	10mg
Vials/Units Received	1

Reagent Information

Cartridge Type	Endotoxin- LAL (Charles River)
Cartridge Lot #	2520301152
Cartridge Calibration Code	5136352622125
Detection Range	0.05 EU/mL – 5.0 EU/mL
Spike Value	0.474 EU/mL
Spike Recovery	115% (Pass)
Test Suitability	Pass

Endotoxin Quantitation Results

Sample ID	Replicates	Measured EU/mL	Specification	Pass/Fail
BPC-157	1, 2	0.07 EU/mL	≤ 0.50 EU/mL	Pass

Interpretation:

Endotoxin level in **BPC-157** was measured using chromogenic LAL assay with detection at 405 nm. The result of **<0.07 EU/mL** indicates that the endotoxin content is **below** the acceptable threshold of 0.5 EU/mL. Sample **PASSES** the endotoxin limit test.



Analyst Signature: _____

Date: 02/17/2026

To verify the authenticity of this report, please contact **BioRegen** at contact@bioregen.com or reference the **Report ID** and **Validation Key** provided above.

Alternatively, **scan the QR code** below to access the official record on our website.

