BioRegen
Independent Testing, Trusted Results

contact@bioregen.com
www.bioregen.com

Report ID: NC-ENDO-04272026-445

Validation Key: NOVCOMP01

Limulus Amebocyte Lysate (LAL) – Report of Analysis

Test Performed	Endosafe - PTS LAL Assay
Test Date	04/27/2026
Analyst	VICTOR
Instrument	Endosafe® nexgen-PTS™ (SN: 25201216)

Sample Information

Company Name	Novera Compounds
Sample Tested	Tesamorelin/Ipamorelin
Sample Lot Number	2X12A013026
Sample Concentration	10mg/2mg
Vials/Units Received	1

Reagent Information

Cartridge Type	Endotoxin- LAL (Charles River)
Cartridge Lot #	2602741388
Cartridge Calibration Code	518334799106
Detection Range	0.05 EU/mL – 5.0 EU/mL
Spike Value	1.19 EU/mL
Spike Recovery	107% (Pass)
Test Suitability	Pass

Endotoxin Quantitation Results

Sample ID	Replicates	Measured EU/mL	Specification	Pass/Fail
Retatrutide	1, 2	<0.05 EU/mL	≤ 0.50 EU/mL	Pass

Interpretation:
Endotoxin level in **Retatrutide** was measured using chromogenic LAL assay with detection at 405 nm. The result of **<0.05 EU/mL** indicates that the endotoxin content is **below** the acceptable threshold of 0.5 EU/mL. Sample **PASSES** the endotoxin limit test.

Analyst Signature: _____
Date: 04/30/2026

To verify the authenticity of this report, please contact **BioRegen** at **contact@bioregen.com** or reference the **Report ID** and **Validation Key** provided above.

Alternatively, **scan the QR code** below to access the official record on our website.